

Operating Manual for EASYBUS - Sensor Modules

EBH - 1R / 1K / 2K

Specification:

Measuring range:	please refer to type plate Standard: 0 - 100 %RH (rec. range of application: 30 - 80 %RH) High humidity: 0 - 100 %RH (rec. range of application: 11 - 90 %RH)
Resolution:	0,1%
Sensor:	Capacitive thin film sensor (unsuitable for dusty and aggressive media)
Display: [optional]	3½-digit LC-display
Linearity:	Standard: ±2 %RH (from 30 to 80 %RH) High humidity: ±2 %RH (from 11 to 90 %RH)
Hysteresis:	Standard ±1,5 %RH (from 30 to 80 %RH) High humidity: ±1,0 %RH (from 11 to 90 %RH)
Nominal temperature:	25°C
Working temperature:	electronic: 0 to 70°C sensor: -30 to 120°C (at option high temperatur)
Storage temperature:	-20 to 70°C
Relative atmospheric humidity:	0 to 98 % r.h. (non-condensing)
Housing:	ABS (IP65)
Sensor tube:	Aluminium, Ø 14 mm, anodized with bronze filter 160 µm
Mounting:	By means of screw thread or mounting holes in housing (accessible after cover has been removed)
Mounting position:	sensor tube pointing vertically downwards, ±90° angle deviation
Mounting distance:	50 x 70mm, max. shaft diameter of mounting screws is 4mm
Electric connection:	2-wire connection, no polarity Via elbow-type plug according to DIN 43650 (IP65) for cables up to 1,5mm ² . EASYBUS -connection at elbow-type plug terminal 1 and 2
Busload:	1.5 EASYBUS -devices
EMC:	The device corresponds to the essential protection ratings established in the Regulations of the Council for the Approximation of Legislation for the member countries regarding electromagnetic compatibility (89/336/EWG). Additional fault: <1%

General installation instructions:

To mount the connection cable the elbow-type plug screw has to be loosened and the coupling insert has to be removed by means of a screw driver at the position indicated (arrow). Pull out connection cable through PG glanding and connect to the loose coupling insert as described in the wiring diagram. Replace loose coupling insert onto the pins at the transmitter housing and turn cover cap with PG glanding in the direction desired till it snaps on. (4 different starting positions at 90° intervals). Re-tighten the screw at the elbow-type plug.

The **EASYBUS**-connection is made with the elbow-type plug terminals 1 and 2.

Safety instructions:

This device has been designed and tested in accordance with the safety regulations for electronic devices.

However, its trouble-free operation and reliability cannot be guaranteed unless the standard safety measures and special safety advises given in this manual will be adhered to when using the device.

1. Trouble-free operation and reliability of the device can only be guaranteed if the device is not subjected to any other climatic conditions than those stated under "Specification".
If the device is transported from a cold to a warm environment condensation may cause in a failure of the function. In such a case make sure the device temperature has adjusted to the ambient temperature before trying a new start-up.
2. General instructions and safety regulations for electric, light and heavy current plants, including domestic safety regulations (e.g. VDE), have to be observed.
3. If device is to be connected to other devices (e.g. via PC) the circuitry has to be designed most carefully. Internal connection in third party devices (e.g. connection GND and earth) may result in not-permissible voltages impairing or destroying the device or another device connected.
4. If there is a risk whatsoever involved in running it, the device has to be switched off immediately and to be marked accordingly to avoid re-starting.
Operator safety may be a risk if:
 - there is visible damage to the device
 - the device is not working as specified
 - the device has been stored under unsuitable conditions for a longer time.

In case of doubt, please return device to manufacturer for repair or maintenance.

